



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT

BREATH ALCOHOL PROGRAM

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN Owensville Police Department 201292	DATE OF INSPECTION 06-2-09
LOCATION OF INSTRUMENT (STREET AND CITY) 109 North Second street Owensville	TIME OF INSPECTION 0424

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE **0424 06/02/09**

☒ SIMULATOR TEMPERATURE (34 °C ± 0.2°C) **34.0 °C**

☒ CALIBRATION CHECK -
Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 .098	TEST 2 .099	TEST 3 .100
--------------------	--------------------	--------------------

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) /	(.15-.19) 0	(Over .19) 0
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

Instrument operating properly within Department of Health and Senior Services guidelines and standards.

RepCo Marketing, Inc. Lot #: 08002 .10 solution exp. date: 10-13-10
bottle #: 0798

INSPECTING OFFICER	
SIGNATURE [Signature]	PRINT NAME JONATHAN Scott GRIFFITH
TYPE II PERMIT NUMBER/EXPIRATION DATE 820009 / 01-09-2010	TELEPHONE NUMBER 573-437-2195

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 08002
EXPIRATION DATE: October 13, 2010 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 08002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed utilizing a gas chromatograph and found to contain 0.1209 gms/dl wt/vol ethyl alcohol in aqueous solution (i.e. ethanol).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of 0.100% +/-3% when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator.

The date of manufacture for this lot number is October 14, 2008. The expiration date for this lot number is October 13, 2010 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marking, Inc.

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
OWENSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201292
06/02/09

ARREST TIME: 04:15
SUBJECT NAME:
RFI/TEST
DOB: 09/09/09 SEX: M
STATE/D.L.: NA/NA
ARRESTING OFFICER:
NA

OFFICER I.D.: NA
TESTING OFFICER:
GRIFFITH/JONATHAN/SCOTT
OFFICER I.D.: 104
PERMIT NUMBER: 820009
EXPIRATION DATE: 01/09/10
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST .000 04:29
INTERNAL STANDARD VERIFIED 04:30
RADIO INTERFERENCE

Operator Signature

[Signature]

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
OWENSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201292
06/02/09

TESTING OFFICER:
GRIFFITH/JONATHAN/SCOTT
OFFICER I.D.: 104
PERMIT NUMBER: 820009
EXPIRATION DATE: 01/09/10
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST .000 04:29
INTERNAL STANDARD VERIFIED 04:29
EXTERNAL STANDARD .098 04:29
BLANK TEST .000 04:30
EXTERNAL STANDARD .099 04:30
BLANK TEST .000 04:31
EXTERNAL STANDARD .100 04:31
BLANK TEST .000 04:32

N = 3

SIM. = .1

AVG. = .099

Operator Signature

[Signature]

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
JENNINGSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201292

06/02/09

04:24

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

HEATERS: 49C

SAMPLE CHAMBER: 49C

FLOW DETECTOR: OKAY

PUMP: OKAY

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

GURPZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*~.,/0123456789:;<=>?ABCD EFG
HIJKLMN O PQRSTU VWXYZ[\]^_`ab cdefgh i j k l m n o
p q r s t u v w x y z { | } ~

Operator Signature

[Signature]

220A-02

State of Missouri DEPARTMENT OF HEALTH

PERMIT TYPE II



JONATHAN S. GRIFFITH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 01/09/08

Number 820009

Expires 01/09/2010

[Signature]
Director of State Public Health Laboratory

[Signature]
Director, Department of Health